

TEST AND CALIBRATION CERTIFICATE

DATE CERTIFICATE N^o.
 LOAD CELL TYPE SERIAL N^o.
 CALIBRATION METHOD CALIBRATION MODE
 BRIDGE RESISTANCE INSULATION RESISTANCE
 (NOMINAL)

CONNECTIONS

+VE SUPPLY	PIN 1 / RED	-VE SUPPLY	PIN 2 / BLUE
+VE OUTPUT	PIN 3 / GREEN	-VE OUTPUT	PIN 4 / YELLOW
	SCREEN	BLACK	

PROOF LOAD

LOAD TONNE	AVERAGE READING mV/V	BSL mV/V	DEVIATION FROM BSL % FRO
0.00	0.0000	0.0000	0.00
40.00	0.3280	0.3273	0.04
80.00	0.6560	0.6546	0.09
120.00	0.9830	0.9819	0.07
160.00	1.3090	1.3091	-0.01
200.00	1.6350	1.6364	-0.09

ZERO OFFSET mV/V CAL. READING mV/V
 CAL. RESISTOR VALUE KΩ CAL. EQUIVALENT

CALIBRATED BY:

WEBMASTER

ALL LOADS APPLIED ARE TRACEABLE TO NATIONAL STANDARDS

Registered Office and Trading Address:

LCM Systems Limited • Unit 15, Newport Business Park • Barry Way • Newport • Isle of Wight • PO30 5GY • UK
 Tel: +44 (0)1983 249264 • Fax: +44 (0)1983 249266 • Email: info@lcmsystems.com

[Http: www.lcmsystems.com](http://www.lcmsystems.com)

Registered in England No. 2057541

TEST AND CALIBRATION CERTIFICATE

DATE CERTIFICATE N^o.
 LOAD CELL TYPE SERIAL N^o.
 CALIBRATION METHOD CALIBRATION MODE
 BRIDGE RESISTANCE INSULATION RESISTANCE
 (NOMINAL)

CONNECTIONS

+VE SUPPLY	PIN 1 / RED	-VE SUPPLY	PIN 2 / BLUE
+VE OUTPUT	PIN 3 / GREEN	-VE OUTPUT	PIN 4 / YELLOW
	SCREEN	BLACK	

PROOF LOAD

LOAD TONNE	AVERAGE READING mV/V	BSL mV/V	DEVIATION FROM BSL % FRO
0.00	0.0000	0.0000	0.00
40.00	0.3330	0.3297	0.20
80.00	0.6620	0.6593	0.16
120.00	0.9900	0.9890	0.06
160.00	1.3170	1.3187	-0.10
200.00	1.6450	1.6483	-0.20

ZERO OFFSET mV/V CAL. READING mV/V
 CAL. RESISTOR VALUE KΩ CAL. EQUIVALENT

CALIBRATED BY:WEBMASTER**ALL LOADS APPLIED ARE TRACEABLE TO NATIONAL STANDARDS***Registered Office and Trading Address:*

LCM Systems Limited • Unit 15, Newport Business Park • Barry Way • Newport • Isle of Wight • PO30 5GY • UK
 Tel: +44 (0)1983 249264 • Fax: +44 (0)1983 249266 • Email: info@lcmsystems.com

Http: www.lcmsystems.com

TEST AND CALIBRATION CERTIFICATE

DATE CERTIFICATE N^o.
 LOAD CELL TYPE SERIAL N^o.
 CALIBRATION METHOD CALIBRATION MODE
 BRIDGE RESISTANCE INSULATION RESISTANCE
 (NOMINAL)

CONNECTIONS

+VE SUPPLY	PIN 1 / RED	-VE SUPPLY	PIN 2 / BLUE
+VE OUTPUT	PIN 3 / GREEN	-VE OUTPUT	PIN 4 / YELLOW
	SCREEN	BLACK	

PROOF LOAD

LOAD TONNE	AVERAGE READING mV/V	BSL mV/V	DEVIATION FROM BSL % FRO
0.00	0.0000	0.0000	0.00
40.00	0.3280	0.3280	0.00
80.00	0.6560	0.6560	0.00
120.00	0.9840	0.9840	0.00
160.00	1.3120	1.3120	0.00
200.00	1.6400	1.6400	0.00

ZERO OFFSET mV/V CAL. READING mV/V
 CAL. RESISTOR VALUE KΩ CAL. EQUIVALENT

CALIBRATED BY:WEBMASTER**ALL LOADS APPLIED ARE TRACEABLE TO NATIONAL STANDARDS***Registered Office and Trading Address:*

LCM Systems Limited • Unit 15, Newport Business Park • Barry Way • Newport • Isle of Wight • PO30 5GY • UK
 Tel: +44 (0)1983 249264 • Fax: +44 (0)1983 249266 • Email: info@lcmsystems.com

Http: www.lcmsystems.com

TEST AND CALIBRATION CERTIFICATE

DATE CERTIFICATE N^o.
 LOAD CELL TYPE SERIAL N^o.
 CALIBRATION METHOD CALIBRATION MODE
 BRIDGE RESISTANCE INSULATION RESISTANCE
 (NOMINAL)

CONNECTIONS

+VE SUPPLY	PIN 1 / RED	-VE SUPPLY	PIN 2 / BLUE
+VE OUTPUT	PIN 3 / GREEN	-VE OUTPUT	PIN 4 / YELLOW
	SCREEN	BLACK	

PROOF LOAD

LOAD TONNE	AVERAGE READING mV/V	BSL mV/V	DEVIATION FROM BSL % FRO
0.00	0.0000	0.0000	0.00
40.00	0.3280	0.3285	-0.03
80.00	0.6570	0.6570	0.00
120.00	0.9860	0.9855	0.03
160.00	1.3145	1.3140	0.03
200.00	1.6425	1.6425	0.00

ZERO OFFSET mV/V CAL. READING mV/V
 CAL. RESISTOR VALUE KΩ CAL. EQUIVALENT

CALIBRATED BY:WEBMASTERALL LOADS APPLIED ARE TRACEABLE TO NATIONAL STANDARDS*Registered Office and Trading Address:*

LCM Systems Limited • Unit 15, Newport Business Park • Barry Way • Newport • Isle of Wight • PO30 5GY • UK
 Tel: +44 (0)1983 249264 • Fax: +44 (0)1983 249266 • Email: info@lcmsystems.com

Http: www.lcmsystems.com