

TEST AND CALIBRATION CERTIFICATE

DATE CERTIFICATE N^o.
 LOAD CELL TYPE SERIAL N^o.
 CALIBRATION METHOD CALIBRATION MODE
 BRIDGE RESISTANCE (NOMINAL) INSULATION RESISTANCE

CONNECTIONS

+VE SUPPLY	RED	-VE SUPPLY	BLUE
+VE OUTPUT	GREEN	-VE OUTPUT	YELLOW
	SCREEN	BLACK	

PROOF LOAD

LOAD TONNE	AVERAGE READING mV/V	BSL mV/V	DEVIATION FROM BSL % FRO
0.00	0.0000	0.0000	0.00
0.20	0.2475	0.2496	-0.17
0.40	0.4965	0.4991	-0.21
0.60	0.7475	0.7487	-0.10
0.80	0.9990	0.9983	0.06
1.00	1.2505	1.2479	0.21

ZERO OFFSET mV/V CAL. READING mV/V
 CAL. RESISTOR VALUE KΩ CAL. EQUIVALENT

CALIBRATED BY:**WEBMASTER****ALL LOADS APPLIED ARE TRACEABLE TO NATIONAL STANDARDS**

Registered Office and Trading Address:

LCM Systems Limited • Unit 15, Newport Business Park • Barry Way • Newport • Isle of Wight • PO30 5GY • UK
 Tel: +44 (0)1983 249264 • Fax: +44 (0)1983 249266 • Email: info@lcmsystems.com

Http: www.lcmsystems.com

Registered in England No. 2057541

TEST AND CALIBRATION CERTIFICATE

DATE CERTIFICATE N^o.
 LOAD CELL TYPE SERIAL N^o.
 CALIBRATION METHOD CALIBRATION MODE
 BRIDGE RESISTANCE INSULATION RESISTANCE
 (NOMINAL)

CONNECTIONS

+VE SUPPLY	RED	-VE SUPPLY	BLUE
+VE OUTPUT	GREEN	-VE OUTPUT	YELLOW
	SCREEN	BLACK	

PROOF LOAD

LOAD TONNE	AVERAGE READING mV/V	BSL mV/V	DEVIATION FROM BSL % FRO
0.00	0.0000	0.0000	0.00
0.20	0.2855	0.2895	-0.28
0.40	0.5765	0.5790	-0.17
0.60	0.8685	0.8685	0.00
0.80	1.1595	1.1580	0.10
1.00	1.4515	1.4475	0.28

ZERO OFFSET mV/V CAL. READING mV/V
 CAL. RESISTOR VALUE KΩ CAL. EQUIVALENT

CALIBRATED BY:**WEBMASTER****ALL LOADS APPLIED ARE TRACEABLE TO NATIONAL STANDARDS***Registered Office and Trading Address:*

LCM Systems Limited • Unit 15, Newport Business Park • Barry Way • Newport • Isle of Wight • PO30 5GY • UK
 Tel: +44 (0)1983 249264 • Fax: +44 (0)1983 249266 • Email: info@lcmsystems.com

Http: www.lcmsystems.com

Registered in England No. 2057541

TEST AND CALIBRATION CERTIFICATE

DATE CERTIFICATE N^o.
 LOAD CELL TYPE SERIAL N^o.
 CALIBRATION METHOD CALIBRATION MODE
 BRIDGE RESISTANCE INSULATION RESISTANCE
 (NOMINAL)

CONNECTIONS

+VE SUPPLY	RED	-VE SUPPLY	BLUE
+VE OUTPUT	GREEN	-VE OUTPUT	YELLOW
	SCREEN	BLACK	

PROOF LOAD

LOAD TONNE	AVERAGE READING mV/V	BSL mV/V	DEVIATION FROM BSL % FRO
0.00	0.0000	0.0000	0.00
0.20	0.2420	0.2441	-0.17
0.40	0.4890	0.4882	0.07
0.60	0.7335	0.7322	0.10
0.80	0.9780	0.9763	0.14
1.00	1.2225	1.2204	0.17

ZERO OFFSET mV/V CAL. READING mV/V CAL. RESISTOR VALUE KΩ CAL. EQUIVALENT **CALIBRATED BY:****WEBMASTER****ALL LOADS APPLIED ARE TRACEABLE TO NATIONAL STANDARDS***Registered Office and Trading Address:*

LCM Systems Limited • Unit 15, Newport Business Park • Barry Way • Newport • Isle of Wight • PO30 5GY • UK
 Tel: +44 (0)1983 249264 • Fax: +44 (0)1983 249266 • Email: info@lcmsystems.com

Http: www.lcmsystems.com

Registered in England No. 2057541

TEST AND CALIBRATION CERTIFICATE

DATE CERTIFICATE N^o.

LOAD CELL TYPE SERIAL N^o.

CALIBRATION METHOD CALIBRATION MODE

BRIDGE RESISTANCE INSULATION RESISTANCE

(NOMINAL)

CONNECTIONS

+VE SUPPLY	RED	-VE SUPPLY	BLUE
+VE OUTPUT	GREEN	-VE OUTPUT	YELLOW
	SCREEN	BLACK	

PROOF LOAD

LOAD TONNE	AVERAGE READING mV/V	BSL mV/V	DEVIATION FROM BSL % FRO
0.00	0.0000	0.0000	0.00
0.20	0.2200	0.2319	-1.03
0.40	0.4535	0.4638	-0.89
0.60	0.6925	0.6957	-0.28
0.80	0.9325	0.9277	0.42
1.00	1.1715	1.1596	1.03

ZERO OFFSET mV/V CAL. READING mV/V CAL. RESISTOR VALUE KΩ CAL. EQUIVALENT CALIBRATED BY:WEBMASTER**ALL LOADS APPLIED ARE TRACEABLE TO NATIONAL STANDARDS***Registered Office and Trading Address:*

LCM Systems Limited • Unit 15, Newport Business Park • Barry Way • Newport • Isle of Wight • PO30 5GY • UK
 Tel: +44 (0)1983 249264 • Fax: +44 (0)1983 249266 • Email: info@lcmssystems.com

Http: www.lcmssystems.com

Registered in England No. 2057541