

**CERTIFICATE OF TEST AND CALIBRATION**

DATE OF TEST:  CERTIFICATE No.:   
TEST LOCATION:  MODE:   
LOAD CELL TYPE:  SERIAL No.:   
NOM. BRIDGE RES.:  BRIDGE INSULATION RES. @ 500 Vdc:   
PROOF LOAD:

CONNECTIONS:

PIN 1 : POSITIVE SUPPLY	PIN 2 : NEGATIVE SUPPLY
PIN 3 : POSITIVE SIGNAL	PIN 4 : NEGATIVE SIGNAL

AMPLIFIER TYPE:  SERIAL No.:

OUTPUT FROM AMPLIFIER IN MILLIAMPS	
LOAD (TONNE)	OUTPUT READING 1
0	7.00
40	7.11
80	10.29
120	13.51
160	16.74
200	19.98

CALIBRATION FIGURE:

CALIBRATION CARRIED OUT BY:

WEBMASTER

ALL LOADS APPLIED ARE TRACEABLE TO NATIONAL STANDARDS