

TEST AND CALIBRATION CERTIFICATE

DATE CERTIFICATE N^o.

LOAD CELL TYPE SERIAL N^o.

CALIBRATION METHOD CALIBRATION MODE

BRIDGE RESISTANCE INSULATION RESISTANCE

(NOMINAL)

CONNECTIONS

+VE SUPPLY	PIN 1	-VE SUPPLY	PIN 2
+VE OUTPUT	PIN 3	-VE OUTPUT	PIN 4

PROOF LOAD

LOAD TONNE	AVERAGE READING mV/V	BSL mV/V	DEVIATION FROM BSL % FRO
0.00	0.0000	0.0000	0.00
40.00	0.2605	0.2662	-0.43
80.00	0.5275	0.5323	-0.36
120.00	0.7970	0.7985	-0.11
160.00	1.0665	1.0647	0.14
200.00	1.3365	1.3308	0.43

ZERO OFFSET mV/V CAL. READING mV/V

CAL. RESISTOR VALUE KΩ CAL. EQUIVALENT

CALIBRATED BY:WEBMASTERALL LOADS APPLIED ARE TRACEABLE TO NATIONAL STANDARDS

Registered Office and Trading Address:

LCM Systems Limited • Unit 15, Newport Business Park • Barry Way • Newport • Isle of Wight • PO30 5GY • UK
 Tel: +44 (0)1983 249264 • Fax: +44 (0)1983 249266 • Email: info@lcm systems.com

[Http: www.lcm systems.com](http://www.lcm systems.com)

Registered in England No. 2057541