

TEST AND CALIBRATION CERTIFICATE

DATE CERTIFICATE N^o.

LOAD CELL TYPE SERIAL N^o.

CALIBRATION METHOD CALIBRATION MODE

BRIDGE RESISTANCE INSULATION RESISTANCE

(NOMINAL)

CONNECTIONS

+VE SUPPLY	PIN 1	-VE SUPPLY	PIN 2
+VE OUTPUT	PIN 3	-VE OUTPUT	PIN 4

PROOF LOAD

LOAD TONNE	AVERAGE READING mV/V	BSL mV/V	DEVIATION FROM BSL % FRO
0.00	0.0000	0.0000	0.00
40.00	0.5075	0.4861	0.88
80.00	0.9925	0.9722	0.84
120.00	1.4690	1.4583	0.44
160.00	1.9420	1.9443	-0.10
200.00	2.4090	2.4304	-0.88

ZERO OFFSET mV/V CAL. READING mV/V CAL. RESISTOR VALUE KΩ CAL. EQUIVALENT CALIBRATED BY:WEBMASTER**ALL LOADS APPLIED ARE TRACEABLE TO NATIONAL STANDARDS**

Registered Office and Trading Address:

LCM Systems Limited • Unit 15, Newport Business Park • Barry Way • Newport • Isle of Wight • PO30 5GY • UK
 Tel: +44 (0)1983 249264 • Fax: +44 (0)1983 249266 • Email: info@lcmsystems.com

Http: www.lcmsystems.com

Registered in England No. 2057541