

**TEST AND CALIBRATION CERTIFICATE**

DATE  CERTIFICATE N<sup>o</sup>.   
LOAD CELL TYPE  SERIAL N<sup>o</sup>.   
CALIBRATION METHOD  CALIBRATION MODE   
BRIDGE RESISTANCE  INSULATION RESISTANCE   
(NOMINAL)

**CONNECTIONS**

+VE SUPPLY	PIN 1	-VE SUPPLY	PIN 2
+VE OUTPUT	PIN 3	-VE OUTPUT	PIN 4

PROOF LOAD 

LOAD TONNE	AVERAGE READING mV/V	BSL mV/V	DEVIATION FROM BSL % FRO
0.00	0.0000	0.0000	0.00
30.00	0.4380	0.4062	1.57
60.00	0.8290	0.8123	0.82
90.00	1.2260	1.2185	0.37
120.00	1.6150	1.6247	-0.48
150.00	1.9990	2.0308	-1.57

ZERO OFFSET mV/V  CAL. READING mV/V   
CAL. RESISTOR VALUE KΩ  CAL. EQUIVALENT

CALIBRATED BY:WEBMASTER**ALL LOADS APPLIED ARE TRACEABLE TO NATIONAL STANDARDS***Registered Office and Trading Address:*LCM Systems Limited • Unit 15, Newport Business Park • Barry Way • Newport • Isle of Wight • PO30 5GY • UK  
Tel: +44 (0)1983 249264 • Fax: +44 (0)1983 249266 • Email: info@lcmsystems.com**Http: [www.lcmsystems.com](http://www.lcmsystems.com)**

Registered in England No. 2057541