

TEST AND CALIBRATION CERTIFICATE

DATE CERTIFICATE N^o.
 LOAD CELL TYPE SERIAL N^o.
 CALIBRATION METHOD CALIBRATION MODE
 BRIDGE RESISTANCE INSULATION RESISTANCE
 (NOMINAL)

CONNECTIONS

+VE SUPPLY	RED	-VE SUPPLY	BLUE
+VE OUTPUT	GREEN	-VE OUTPUT	YELLOW
	SCREEN	BLACK	

PROOF LOAD

LOAD KG	AVERAGE READING mV/V	BSL mV/V	DEVIATION FROM BSL % FRO
0.00	0.0000	0.0000	0.00
70.00	0.0834	0.0845	-0.25
140.00	0.1678	0.1689	-0.27
210.00	0.2520	0.2534	-0.32
280.00	0.3379	0.3378	0.01
350.00	0.4236	0.4223	0.32

ZERO OFFSET mV/V CAL. READING mV/V CAL. RESISTOR VALUE KΩ CAL. EQUIVALENT **CALIBRATED BY:****WEBMASTER****ALL LOADS APPLIED ARE TRACEABLE TO NATIONAL STANDARDS**

TEST AND CALIBRATION CERTIFICATE

DATE CERTIFICATE N^o.
 LOAD CELL TYPE SERIAL N^o.
 CALIBRATION METHOD CALIBRATION MODE
 BRIDGE RESISTANCE INSULATION RESISTANCE
 (NOMINAL)

CONNECTIONS

+VE SUPPLY	RED	-VE SUPPLY	BLUE
+VE OUTPUT	GREEN	-VE OUTPUT	YELLOW
	SCREEN	BLACK	

PROOF LOAD

LOAD KG	AVERAGE READING mV/V	BSL mV/V	DEVIATION FROM BSL % FRO
0.00	0.0000	0.0000	0.00
70.00	0.0900	0.0896	0.09
140.00	0.1791	0.1791	-0.01
210.00	0.2683	0.2687	-0.09
280.00	0.3579	0.3582	-0.07
350.00	0.4477	0.4477	-0.01

ZERO OFFSET mV/V CAL. READING mV/V
 CAL. RESISTOR VALUE KΩ CAL. EQUIVALENT

CALIBRATED BY:WEBMASTER**ALL LOADS APPLIED ARE TRACEABLE TO NATIONAL STANDARDS***Registered Office and Trading Address:*

LCM Systems Limited • Unit 15, Newport Business Park • Barry Way • Newport • Isle of Wight • PO30 5GY • UK
 Tel: +44 (0)1983 249264 • Fax: +44 (0)1983 249266 • Email: info@lcm systems.com

Http: www.lcm systems.com

Registered in England No. 2057541

TEST AND CALIBRATION CERTIFICATE

DATE CERTIFICATE N^o.
 LOAD CELL TYPE SERIAL N^o.
 CALIBRATION METHOD CALIBRATION MODE
 BRIDGE RESISTANCE INSULATION RESISTANCE
 (NOMINAL)

CONNECTIONS

+VE SUPPLY	RED	-VE SUPPLY	BLUE
+VE OUTPUT	GREEN	-VE OUTPUT	YELLOW
	SCREEN	BLACK	

PROOF LOAD

LOAD KG	AVERAGE READING mV/V	BSL mV/V	DEVIATION FROM BSL % FRO
0.00	0.0000	0.0000	0.00
70.00	0.0871	0.0870	0.02
140.00	0.1739	0.1740	-0.03
210.00	0.2612	0.2611	0.02
280.00	0.3482	0.3481	0.03
350.00	0.4350	0.4351	-0.02

ZERO OFFSET mV/V CAL. READING mV/V
 CAL. RESISTOR VALUE KΩ CAL. EQUIVALENT

CALIBRATED BY:**WEBMASTER****ALL LOADS APPLIED ARE TRACEABLE TO NATIONAL STANDARDS***Registered Office and Trading Address:*

LCM Systems Limited • Unit 15, Newport Business Park • Barry Way • Newport • Isle of Wight • PO30 5GY • UK
 Tel: +44 (0)1983 249264 • Fax: +44 (0)1983 249266 • Email: info@lcm systems.com

Http: www.lcm systems.com

TEST AND CALIBRATION CERTIFICATE

DATE CERTIFICATE N^o.
 LOAD CELL TYPE SERIAL N^o.
 CALIBRATION METHOD CALIBRATION MODE
 BRIDGE RESISTANCE INSULATION RESISTANCE
 (NOMINAL)

CONNECTIONS

+VE SUPPLY	RED	-VE SUPPLY	BLUE
+VE OUTPUT	GREEN	-VE OUTPUT	YELLOW
	SCREEN	BLACK	

PROOF LOAD

LOAD KG	AVERAGE READING mV/V	BSL mV/V	DEVIATION FROM BSL % FRO
0.00	0.0000	0.0000	0.00
70.00	0.0877	0.0896	-0.42
140.00	0.1770	0.1792	-0.50
210.00	0.2672	0.2688	-0.35
280.00	0.3585	0.3584	0.03
350.00	0.4502	0.4480	0.50

ZERO OFFSET mV/V CAL. READING mV/V
 CAL. RESISTOR VALUE KΩ CAL. EQUIVALENT

CALIBRATED BY:WEBMASTERALL LOADS APPLIED ARE TRACEABLE TO NATIONAL STANDARDS*Registered Office and Trading Address:*

LCM Systems Limited • Unit 15, Newport Business Park • Barry Way • Newport • Isle of Wight • PO30 5GY • UK
 Tel: +44 (0)1983 249264 • Fax: +44 (0)1983 249266 • Email: info@lcmssystems.com

[Http: www.lcmssystems.com](http://www.lcmssystems.com)