

CERTIFICATE OF TEST AND CALIBRATION

DATE OF TEST: CERTIFICATE No.:
TEST LOCATION: MODE:
LOAD CELL TYPE: SERIAL No.:
NOM. BRIDGE RES.: BRIDGE INSULATION RES. @ 500 Vdc:
PROOF LOAD:

CONNECTIONS:

<input type="text" value="PIN 1 : POSITIVE SUPPLY"/>	<input type="text" value="PIN 2 : NEGATIVE SUPPLY"/>
<input type="text" value="PIN 3 : POSITIVE SIGNAL"/>	<input type="text" value="PIN 4 : NEGATIVE SIGNAL"/>

DISPLAY TYPE: SERIAL No.:

CHANNEL		
LOAD (TONNE)	DISPLAY READING 1	DISPLAY READING 2
0	0.0	0.0
50	53.2	53.2
100	103.0	103.2
150	152.0	151.8
200	200.0	200.0
250	247.8	247.8

CALIBRATION FIGURE:

CALIBRATION CARRIED OUT BY:

LCM Systems Ltd.

ALL LOADS APPLIED ARE TRACEABLE TO NATIONAL STANDARDS

CERTIFICATE OF TEST AND CALIBRATION

DATE OF TEST: CERTIFICATE No.:
TEST LOCATION: MODE:
LOAD CELL TYPE: SERIAL No.:
NOM. BRIDGE RES.: BRIDGE INSULATION RES. @ 500 Vdc:
PROOF LOAD:

CONNECTIONS:

<input type="text" value="PIN 1 : POSITIVE SUPPLY"/>	<input type="text" value="PIN 2 : NEGATIVE SUPPLY"/>
<input type="text" value="PIN 3 : POSITIVE SIGNAL"/>	<input type="text" value="PIN 4 : NEGATIVE SIGNAL"/>

DISPLAY TYPE: SERIAL No.:

CHANNEL		
LOAD (TONNE)	DISPLAY READING 1	DISPLAY READING 2
0	0.0	0.0
50	52.6	52.8
100	102.8	103.0
150	151.6	151.8
200	199.8	200.0
250	247.6	247.6

CALIBRATION FIGURE:

CALIBRATION CARRIED OUT BY:

LCM Systems Ltd.

ALL LOADS APPLIED ARE TRACEABLE TO NATIONAL STANDARDS