

CERTIFICATE OF TEST AND CALIBRATION

DATE OF TEST: CERTIFICATE No.:
TEST LOCATION: MODE:
LOAD CELL TYPE: SERIAL No.:
NOM. BRIDGE RES.: BRIDGE INSULATION RES. @ 50 Vdc:
PROOF LOAD:

CONNECTIONS:

PIN 1 : POSITIVE SUPPLY	PIN 2 : NEGATIVE SUPPLY
PIN 3 : POSITIVE SIGNAL	PIN 4 : NEGATIVE SIGNAL

OUTPUT FROM INTERNAL ICA1 AMPLIFIER		
LOAD (KN)	OUTPUT (Vdc) READING 1	OUTPUT (Vdc) READING 2
0	0.10	0.10
70	2.14	2.13
140	4.15	4.15
210	6.11	6.11
280	8.00	8.00
350	9.88	9.86

CALIBRATION FIGURE:

CALIBRATION CARRIED OUT BY:

WEBMASTER

ALL LOADS APPLIED ARE TRACEABLE TO NATIONAL STANDARDS